

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

### Measurement Uncertainty: Summary Table for Phadia Assays

Measurement Uncertainty (MU) is calculated using internal quality control (IQC)

The raw data and calculations can be found at the following location: <\\xggc-fsrv-04\GGC Biochemistry\Immunology\1IMM&NI\Quality\Uncertainty of Measurement>

Analyte	<b>IgA anti-TTG antibodies on Phadia 2500</b>	
	Low IQC	High IQC
Mean (x)	16.2	67.0
Number of measurements (n)	81	162
Estimated Standard Deviation (s)	1.37	4.57
Coefficient of Variance (%CV)	8.47%	6.82%
Coverage factor (k) to define a confidence level of 95%	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 16.9%</b>	<b>± 13.6%</b>
Uncertainty of measurement <b>example</b> (using IQC)	16.2 ± 2.74 U/mL (13.4 – 18.9 U/mL)	67.0 ± 9.14 U/mL (57.8 – 76.1 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>IgG anti-TTG antibodies on Phadia 2500</b>
	IgG TTG IQC
Mean (x)	33.6
Number of measurements (n)	65
Estimated Standard Deviation (s)	3.22
Coefficient of Variance (%CV)	9.59%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 19.2%</b>
Uncertainty of measurement <b>example</b> (using IQC)	33.6 ± 6.44 U/mL (27.1 – 40.0 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-MPO antibodies on Phadia</b>	
	MPO IQC on <u>Phadia 2500</u>	MPO IQC on <u>Phadia 250</u>
Mean (x)	35.7	36.9
Number of measurements (n)	164	13
Estimated Standard Deviation (s)	2.91	2.84
Coefficient of Variance (%CV)	8.15%	7.70%
Coverage factor (k) to define a confidence level of 95%	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 16.3%</b>	<b>± 15.4%</b>
Uncertainty of measurement <b>example</b> (using IQC)	35.7 ± 5.8 IU/mL (29.9 – 41.5 IU/mL)	36.9 ± 5.7 IU/mL (31.2 – 42.6 IU/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-PR3 antibodies on Phadia</b>	
	<u>PR3 IQC on Phadia 2500</u>	<u>PR3 IQC on Phadia 250</u>
Mean (x)	29.3	23.7
Number of measurements (n)	225	18
Estimated Standard Deviation (s)	3.60	3.21
Coefficient of Variance (%CV)	12.32%	13.55%
Coverage factor (k) to define a confidence level of 95%	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 24.6%</b>	<b>± 27.1%</b>
Uncertainty of measurement <b>example</b> (using IQC)	29.3 ± 7.21 IU/mL (22.0 – 36.5 IU/mL)	23.7 ± 6.43 IU/mL (17.3 – 30.1 IU/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-GBM antibodies on Phadia</b>	
	GBM IQC on <u>Phadia 2500</u>	GBM IQC on <u>Phadia 250</u>
Mean (x)	66.3	62.8
Number of measurements (n)	78	4*
Estimated Standard Deviation (s)	6.39	2.22
Coefficient of Variance (%CV)	9.64%	3.53%
Coverage factor (k) to define a confidence level of 95%	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 19.3%</b>	<b>± 7.1%</b>
Uncertainty of measurement <b>example</b> (using IQC)	66.3 ± 12.78 U/mL (53.5 – 79.1 U/mL)	62.8 ± 4.43 U/mL (58.3 – 67.2 U/mL)

\*Note: small number of measurements available for MU (includes monthly analyser comparison data) –250D only used as a backup analyser

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-CCP antibodies on Phadia 2500</b>
	CCP IQC
Mean (x)	73.8
Number of measurements (n)	301
Estimated Standard Deviation (s)	7.25
Coefficient of Variance (%CV)	9.82%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 19.6%</b>
Uncertainty of measurement <b>example</b> (using IQC)	73.8 ± 14.5 U/mL (59.3 – 88.3 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-dsDNA antibodies on Phadia 2500</b>	
	dsDNA Low IQC	dsDNA High IQC
Mean (x)	17.9	75.8
Number of measurements (n)	174	51
Estimated Standard Deviation (s)	1.80	8.67
Coefficient of Variance (%CV)	10.08%	11.44%
Coverage factor (k) to define a confidence level of 95%	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 20.2%</b>	<b>± 22.9%</b>
Uncertainty of measurement <b>example</b> (using IQC)	17.9 ± 3.6 IU/mL (14.3 – 21.5 IU/mL)	75.8 ± 17.3 IU/mL (58.5 – 93.1 IU/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-ENA antibodies on Phadia 250 (ENA screen)</b>
	ENA screen (Symphony <sup>S</sup> ) IQC
Mean (x)	3.7
Number of measurements (n)	102
Estimated Standard Deviation (s)	0.39
Coefficient of Variance (%CV)	10.45%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>±20.9%</b>
Uncertainty of measurement <b>example</b> (using IQC)	3.7 ± 0.8 Ratio (2.9 – 4.5 Ratio)



<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-Ro52 antibodies on Phadia 2500</b>
	Ro52 IQC
Mean (x)	28.8
Number of measurements (n)	40
Estimated Standard Deviation (s)	2.08
Coefficient of Variance (%CV)	7.2%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>±14.4 %</b>
Uncertainty of measurement <b>example</b> (using IQC)	28.8 ± 4.16 U/mL (24.6 – 33.0 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-Ro60 antibodies on Phadia 2500</b>
	Ro60 IQC
Mean (x)	73.1
Number of measurements (n)	159
Estimated Standard Deviation (s)	6.58
Coefficient of Variance (%CV)	8.99%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 18.0%</b>
Uncertainty of measurement <b>example</b> (using IQC)	73.1 ± 13.15 U/mL (60.0 – 86.3 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-La antibodies on Phadia 2500</b>
	La IQC
Mean (x)	55.0
Number of measurements (n)	41
Estimated Standard Deviation (s)	3.77
Coefficient of Variance (%CV)	6.86%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 13.7%</b>
Uncertainty of measurement <b>example</b> (using IQC)	55.0 ± 7.55 U/mL (47.5 – 62.6 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	Anti-SmD antibodies on Phadia 2500
	SmD IQC
Mean (x)	17.7
Number of measurements (n)	25
Estimated Standard Deviation (s)	1.89
Coefficient of Variance (%CV)	10.67%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 21.3%</b>
Uncertainty of measurement <b>example</b> (using IQC)	17.7 ± 3.77 U/mL (13.9 – 21.5 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-U1RNP antibodies on Phadia 2500</b>
	U1-RNP IQC
Mean (x)	15.5
Number of measurements (n)	41
Estimated Standard Deviation (s)	2.46
Coefficient of Variance (%CV)	15.86%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 31.7%</b>
Uncertainty of measurement <b>example</b> (using IQC)	15.5 ± 4.92 U/mL (10.6 – 20.4 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-Jo-1 antibodies on Phadia 2500</b>
	Jo-1 IQC
Mean (x)	34.1
Number of measurements (n)	81
Estimated Standard Deviation (s)	3.93
Coefficient of Variance (%CV)	11.55%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 23.1%</b>
Uncertainty of measurement <b>example</b> (using IQC)	34.1 ± 7.87 U/mL (26.2 – 42.0 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-Scl-70 antibodies on Phadia 2500</b>
	Scl-70 IQC
Mean (x)	37.1
Number of measurements (n)	76
Estimated Standard Deviation (s)	4.26
Coefficient of Variance (%CV)	11.51%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 23.0%</b>
Uncertainty of measurement <b>example</b> (using IQC)	37.1 ± 8.53 U/mL (28.6 – 45.6 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-Ribosomal P antibodies on Phadia 2500</b>
	Ribosomal P IQC
Mean (x)	23.2
Number of measurements (n)	76
Estimated Standard Deviation (s)	3.26
Coefficient of Variance (%CV)	14.04%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 28.1%</b>
Uncertainty of measurement <b>example</b> (using IQC)	23.2 ± 6.52 U/mL (16.7 – 29.7 U/mL)



<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Anti-Centromere B antibodies on Phadia 2500</b>
	Centromere B IQC
Mean (x)	61.0
Number of measurements (n)	85
Estimated Standard Deviation (s)	5.10
Coefficient of Variance (%CV)	8.37%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 16.7%</b>
Uncertainty of measurement <b>example</b> (using IQC)	61.0 ± 10.21 U/mL (50.8 – 71.2 U/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	Total IgE on Phadia 2500	
	Low IQC	High IQC
Mean (x)	146.0	2367.7
Number of measurements (n)	199	152
Estimated Standard Deviation (s)	10.0	167.12
Coefficient of Variance (%CV)	6.85%	7.06%
Coverage factor (k) to define a confidence level of 95%	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 13.7%</b>	<b>± 14.1%</b>
Uncertainty of measurement <b>example</b> (using IQC)	146.0 ± 20.0 kU/L (126– 166 kU/L)	2367.7 ± 334.2 kU/L (2033.5 – 2702.0 kU/L)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	Allergen Specific IgE on Phadia 2500		
	Low IQC	Medium IQC	High IQC
Mean (x)	0.47	6.2	22.0
Number of measurements (n)	160	155	182
Estimated Standard Deviation (s)	0.05	0.59	2.29
Coefficient of Variance (%CV)	9.74%	9.55%	10.38%
Coverage factor (k) to define a confidence level of 95%	2	2	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 19.5%</b>	<b>± 19.1%</b>	<b>± 20.8%</b>
Uncertainty of measurement <b>example</b> (using IQC)	0.47 ± 0.091 kAU/mL (0.37 – 0.56 kAU/mL)	6.2 ± 1.19 kAU/mL (5.0 – 7.4 kAU/mL)	22.0 ± 4.57 kAU/mL (17.4 – 26.6 kAU/mL)

<b>NHS Greater Glasgow &amp; Clyde Immunology and Neuroimmunology</b>		
<b>QF_19</b>	<b>Measurement Uncertainty: Summary Table for Phadia Assays</b>	<b>Version: 4</b>
<b>Author: Lauren Hennessy</b>	<b>Authoriser: Moira Thomas / Sylvia Arthur</b>	<b>Date of Issue: 19/07/24</b>

Analyte	<b>Tryptase on Phadia 250</b>
	Tryptase IQC
Mean (x)	31.4
Number of measurements (n)	34
Estimated Standard Deviation (s)	1.83
Coefficient of Variance (%CV)	5.82%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 11.6%</b>
Uncertainty of measurement <b>example</b> (using IQC)	31.4 ± 3.66 µg/L  (27.7 – 35.1 µg/L)

NHS Greater Glasgow & Clyde Immunology and Neuroimmunology		
QF_19	Measurement Uncertainty: Summary Table for Phadia Assays	Version: 4
Author: Lauren Hennessy	Authoriser: Moira Thomas / Sylvia Arthur	Date of Issue: 19/07/24

Analyte	Specific IgG on Phadia 250
	Aspergillus IgG IQC
Mean (x)	74.8
Number of measurements (n)	143
Estimated Standard Deviation (s)	8.40
Coefficient of Variance (%CV)	11.22%
Coverage factor (k) to define a confidence level of 95%	2
<b>Relative standard expanded uncertainty (U)</b>	<b>± 22.4%</b>
Uncertainty of measurement <b>example</b> (using IQC)	74.8 ± 16.79 mgA/L (58.0 – 91.6 mgA/L)

For the details of the calculation and MU protocol please refer to document QP\_5: Measurement Uncertainty Protocol.